




America

CERTIFICATE

No. U8 109610 0010 Rev. 00

Holder of Certificate: LPKF Laser & Electronics d.o.o.
 Polica 33
 4202 Naklo
 SLOVENIA

Certification Mark:



C US

Product: Laboratory Equipment

Model(s): LPKF ProtoLaser H4

Parameters:

Rated voltage:	115/230 VAC
Rated frequency:	50/60 Hz
Rated power:	0.5 kVA
Protection Class:	I
Laser Class:	1

Tested according to: UL 61010-1:2012/R:2019-07
 CSA C22.2 No. 61010-1:2012/A1:2018-11
 CSA E60825-1:2015

This product was voluntarily tested to the relevant safety requirements referenced on this certificate. It can be marked with the certification mark above. The mark must not be altered in any way. This product certification system operated by TÜV SÜD America Inc. most closely resembles system 3 as defined in ISO/IEC 17067. Certification is based on the TÜV SÜD "Testing and Certification Regulations". TÜV SÜD America Inc. is an OSHA recognized NRTL for USA and a Standards Council of Canada ISO/IEC 17065 accredited Certification body for Canada.

Test report no.: 713254199

Date, 2022-07-27

(Benedikt Pulver)