




America

CERTIFICATE

No. U8 109610 0008 Rev. 01

Holder of Certificate: LPKF Laser & Electronics d.o.o.
Polica 33
4202 Naklo
SLOVENIA

Certification Mark:



Product: Laboratory equipment
(PCB Prototyping device)

Model(s): ProtoMat S104, ProtoMat S64

Parameters:

Rated voltage:	100-240 VAC
Rated frequency:	50/60 Hz
Rated power:	250 W
Protection Class:	I

Tested according to: UL 61010-1:2012/R:2024-11
CSA C22.2 No. 61010-1:2012/U4:2024-11

This product was voluntarily tested to the relevant safety requirements referenced on this certificate. It can be marked with the certification mark above. The mark must not be altered in any way. This product certification system operated by TÜV SÜD America Inc. most closely resembles system 3 as defined in ISO/IEC 17067. Certification is based on the TÜV SÜD "Testing, Certification, Validation and Verification Regulations (TCVVR)". TÜV SÜD America Inc. is an OSHA recognized NRTL for USA and a Standards Council of Canada ISO/IEC 17065 accredited Certification body for Canada.

Test report no.: 713383988

Date, 2025-12-10

Siemon

(Thorsten Siemon)